

AFPC

Sample

07-2009B

Grade

21-0-0

	AOAC Ref.	Method #	# of Labs.	Grand Median	Std Dev
AMMONIACAL NITROGEN					
Ammoniacal Nitrogen, Other		001.99	10	21.03	0.11
Method Group 001.XX PCT			10	21.03	0.14
TOTAL NITROGEN					
Total Nitrogen, Combustion	993.13	010.60	7	21.23	0.12
Total Nitrogen, Other		010.99	1	21.08	0.00
Method Group 010.XX PCT			8	21.18	0.15
FREE WATER					
Free Water, Vacuum Oven	965.08B	060.00	3	0.09	0.05
Free Water, Other		060.99	6	0.11	0.07
Method Group 060.XX PCT			9	0.10	0.11
TOTAL SULFUR (S)					
Sulfur, Other		145.99	2	25.17	0.52
Method Group 145.XX PCT			2	25.17	0.63